

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000913  
15:17 07/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:57  
STERILIZING 5:16  
136.0 C 227kPa 5:16  
135.9 C 229kPa 5:46  
136.3 C 231kPa 6:16  
136.2 C 231kPa 6:46  
136.1 C 231kPa 7:16  
136.2 C 230kPa 7:46  
136.2 C 231kPa 8:16  
136.1 C 231kPa 8:46  
136.1 C 231kPa 8:46  
VENTING 8:46  
AIR DRYING 9:45  
CYCLE COMPLETE 14:34  
-----

Digital Signature #  
AB222E09B3668E3D  
-----