
Midmark M11 - v1.0.0
Total Cycles: 3706

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2023 01:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.2 F 0.0
2:00 144.9 F 2.3
4:00 159.7 F 5.3
6:00 181.4 F 8.9
8:00 200.1 F 13.1
10:00 214.4 F 6.3
12:00 225.0 F 10.7
14:00 229.8 F 7.5
16:00 240.2 F 12.0
18:00 238.4 F 10.1
20:00 250.9 F 16.0
22:00 261.5 F 22.1
24:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.7
1:00 271.7 F 29.1
2:00 272.5 F 29.4
3:00 272.0 F 29.0
4:00 271.1 F 28.6

Min 271.1 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 197.7 F 0.0
10:00 199.9 F 0.0
15:00 201.8 F 0.0
20:00 204.3 F 0.0
25:00 208.8 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:03:13
CYCLE COMPLETE

08/30/2023 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____