
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002021
10:22 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:02
STERILIZING 3:03
136.0 C 221kPa 3:03
136.5 C 226kPa 3:33
136.1 C 222kPa 4:03
136.0 C 222kPa 4:33
136.4 C 225kPa 5:03
136.3 C 225kPa 5:33
136.3 C 225kPa 6:03
136.4 C 226kPa 6:33
136.4 C 226kPa 6:33
VENTING 6:33
AIR DRYING 7:24
CYCLE COMPLETE 37:25

Digital Signature #
ADE7932F7D9408B7
