

-----  
Midmark M11 - v1.0.5  
Total Cycles: 746

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2022 02:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.9 F 0.0  
2:00 151.9 F 2.7  
4:00 174.6 F 6.7  
6:00 196.2 F 11.7  
8:00 216.9 F 6.1  
10:00 231.6 F 11.5  
12:00 233.6 F 9.0  
14:00 239.1 F 10.1  
16:00 247.9 F 14.5  
18:00 260.1 F 21.4  
20:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.6 F 29.6  
2:00 271.7 F 28.8  
3:00 272.1 F 29.6  
4:00 272.5 F 29.4

Min 271.3 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.1 F 0.0  
5:00 221.7 F 0.0  
10:00 229.5 F 0.0  
15:00 221.4 F 0.0  
20:00 227.9 F 0.0  
25:00 225.8 F 0.0  
30:00 223.2 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 20:40  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:10  
CYCLE COMPLETE

02/01/2022 03:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_