
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000059
11:55 26/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:13
STERILIZING 4:35
136.4 C 225kPa 4:35
136.2 C 224kPa 5:05
136.3 C 225kPa 5:35
136.4 C 225kPa 6:05
136.2 C 224kPa 6:35
136.3 C 223kPa 7:05
136.4 C 222kPa 7:35
136.2 C 223kPa 8:05
136.2 C 223kPa 8:05
VENTING 8:05
AIR DRYING 8:57
CYCLE COMPLETE 9:11

Digital Signature #
088EC3A864AC2F4F
