
Midmark M11 - v1.0.0
Total Cycles: 4895

Name: RITTER 1
Sterilizer ID: V2178198
09/11/2024 05:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 180.0 F 0.7
4:00 196.0 F 4.7
6:00 218.0 F 10.0
8:00 221.5 F 5.5
10:00 238.9 F 11.4
12:00 235.9 F 8.9
14:00 238.7 F 9.1
16:00 249.3 F 14.7
18:00 261.7 F 21.8
20:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.7 F 28.9
2:00 272.0 F 28.1
3:00 272.7 F 29.0
4:00 272.2 F 28.3

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.6 F 0.0
5:00 233.5 F 0.0
10:00 200.6 F 0.0
15:00 199.7 F 0.0
20:00 230.2 F 0.0
25:00 230.7 F 0.0
30:00 234.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

09/11/2024 06:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____