

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2790

Name: RITTER2  
Sterilizer ID: V2313228  
08/18/2023 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.6 F 0.0  
2:00 116.5 F 0.0  
4:00 134.2 F 0.0  
6:00 171.3 F 0.0  
8:00 205.3 F 0.0  
10:00 209.5 F 0.0  
12:00 210.3 F 0.0  
14:00 211.2 F 0.0  
16:00 211.4 F 0.0  
18:00 211.4 F 0.0  
20:00 211.4 F 0.0  
22:00 223.9 F 4.1  
24:00 235.5 F 8.6  
26:00 246.0 F 13.4  
28:00 243.8 F 12.3  
30:00 242.8 F 11.9  
32:00 242.4 F 11.7  
34:00 255.9 F 18.7  
36:00 266.7 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 271.8 F 28.8  
2:00 272.1 F 29.0  
3:00 272.7 F 29.4  
4:00 272.8 F 29.4

Min 271.7 F 28.5  
Max 272.9 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 251.2 F 0.0  
10:00 220.2 F 0.0  
15:00 195.9 F 0.0  
20:00 192.9 F 0.0  
25:00 192.0 F 0.0  
30:00 196.7 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 37:53  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:16:08  
CYCLE COMPLETE

08/18/2023 03:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_