
Midmark M11 - v1.0.0
Total Cycles: 1593

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2022 05:42 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.0 F 0.0
2:00 169.6 F 3.5
4:00 181.2 F 7.8
6:00 212.9 F 13.2
8:00 220.9 F 7.4
10:00 234.8 F 12.6
12:00 236.2 F 9.7
14:00 237.1 F 8.7
16:00 249.0 F 14.8
18:00 261.5 F 21.8
20:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 28.9
2:00 271.7 F 28.6
3:00 272.4 F 28.9
4:00 271.5 F 28.3

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 229.3 F 0.0
10:00 233.8 F 0.0
15:00 231.7 F 0.0
20:00 228.9 F 0.0
25:00 228.8 F 0.0
30:00 228.9 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:15
CYCLE COMPLETE

02/02/2022 06:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____