

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3640

Name: RITTER 1  
Sterilizer ID: V2178198  
08/14/2023 07:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.4 F 0.0  
2:00 69.1 F 0.0  
4:00 79.8 F 0.0  
6:00 107.3 F 0.5  
8:00 141.1 F 2.8  
10:00 159.6 F 5.4  
12:00 182.8 F 8.7  
14:00 196.8 F 12.2  
16:00 213.9 F 3.7  
18:00 220.8 F 7.0  
20:00 228.4 F 10.3  
22:00 236.3 F 9.9  
24:00 237.9 F 10.1  
26:00 238.4 F 9.2  
28:00 246.0 F 13.2  
30:00 257.3 F 19.3  
32:00 267.7 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.0 F 28.4  
2:00 272.4 F 28.8  
3:00 272.4 F 29.1  
4:00 271.5 F 28.3  
  
Min 271.3 F 28.2  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.8 F 0.0  
5:00 224.8 F 0.0  
10:00 217.9 F 0.0  
15:00 219.2 F 0.0  
20:00 228.9 F 0.0  
25:00 236.0 F 0.0  
30:00 227.1 F 0.0  
  
Drying Complete

FILLING: 0:48  
HEATING: 33:18  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:11:19  
CYCLE COMPLETE

08/14/2023 08:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_