
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006248
10:00 06/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:34
PRESSURIZING 1:42
STERILIZING 2:34
136.0 C 220kPa 2:34
135.9 C 221kPa 3:04
136.4 C 223kPa 3:34
136.4 C 221kPa 4:04
136.4 C 222kPa 4:34
136.3 C 223kPa 5:04
136.4 C 223kPa 5:34
136.4 C 223kPa 6:04
136.4 C 223kPa 6:04
VENTING 6:04
AIR DRYING 6:56
CYCLE COMPLETE 11:01

Digital Signature #
C0F6C708655A19EA
