
Midmark M11 - v1.0.5
Total Cycles: 3068

Name: RITTER2
Sterilizer ID: V2313228
11/01/2023 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 81.7 F 0.0
4:00 109.4 F 0.3
6:00 104.9 F 1.9
8:00 139.8 F 4.7
10:00 168.1 F 7.9
12:00 217.2 F 11.5
14:00 216.4 F 4.9
16:00 226.8 F 9.2
18:00 239.8 F 13.5
20:00 235.6 F 9.5
22:00 238.1 F 9.8
24:00 244.8 F 13.0
26:00 256.8 F 19.4
28:00 267.7 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 271.9 F 29.6
2:00 272.2 F 29.8
3:00 272.4 F 29.9
4:00 272.3 F 29.9

Min 271.6 F 29.2
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 232.2 F 0.0
10:00 230.1 F 0.0
15:00 189.0 F 0.0
20:00 184.1 F 0.0
25:00 180.9 F 0.0
30:00 178.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:56
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:07:57
CYCLE COMPLETE

11/01/2023 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____