
Midmark M11 - v1.0.0
Total Cycles: 1339

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2021 10:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.2 F 0.0
2:00 190.9 F 2.1
4:00 207.7 F 6.3
6:00 219.6 F 11.7
8:00 226.0 F 6.4
10:00 237.2 F 11.4
12:00 236.1 F 8.5
14:00 242.3 F 10.9
16:00 246.8 F 13.4
18:00 258.8 F 19.9
20:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.5
2:00 272.2 F 28.8
3:00 271.8 F 28.2
4:00 272.6 F 28.9

Min 271.4 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.2 F 0.0
5:00 208.4 F 0.0
10:00 213.2 F 0.0
15:00 216.4 F 0.0
20:00 214.2 F 0.0
25:00 200.7 F 0.0
30:00 194.4 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

11/08/2021 11:56 AM
ACCEPTED: X
REJECTED:
Operator: U01