
Midmark M11 - v1.0.0
Total Cycles: 2427

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2022 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.5 F 0.0
2:00 168.7 F 3.1
4:00 182.5 F 7.9
6:00 204.7 F 13.0
8:00 223.2 F 7.0
10:00 237.4 F 12.1
12:00 234.8 F 9.0
14:00 238.2 F 9.6
16:00 247.0 F 13.9
18:00 259.6 F 20.8
20:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.3 F 28.9
2:00 272.7 F 29.4
3:00 271.9 F 28.5
4:00 272.8 F 29.2

Min 271.6 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.3 F 0.0
5:00 228.3 F 0.0
10:00 230.7 F 0.0
15:00 229.2 F 0.0
20:00 226.1 F 0.0
25:00 219.6 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

10/13/2022 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____