
Midmark M11 - v1.0.5
Total Cycles: 3506

Name: RITTER2
Sterilizer ID: V2313228
04/08/2024 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.2 F 0.0
2:00 165.8 F 2.8
4:00 184.1 F 6.9
6:00 205.5 F 11.9
8:00 218.4 F 6.5
10:00 231.8 F 11.9
12:00 235.1 F 9.7
14:00 236.9 F 9.1
16:00 249.2 F 15.4
18:00 261.1 F 22.2
20:00 272.3 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.5 F 29.5
2:00 272.7 F 29.9
3:00 271.3 F 29.5
4:00 272.3 F 29.6

Min 271.3 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 240.6 F 0.0
10:00 194.1 F 0.0
15:00 188.0 F 0.0
20:00 183.1 F 0.0
25:00 180.9 F 0.0
30:00 179.0 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

04/08/2024 04:45 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____