

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2446

Name: RITTER2  
Sterilizer ID: V2313228  
05/31/2023 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.0 F 0.0  
2:00 202.6 F 3.8  
4:00 201.6 F 9.0  
6:00 221.8 F 8.5  
8:00 231.5 F 10.5  
10:00 232.9 F 8.1  
12:00 240.2 F 10.5  
14:00 247.4 F 14.4  
16:00 260.4 F 21.6  
18:00 271.8 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.4 F 29.8  
2:00 272.1 F 29.1  
3:00 272.1 F 29.6  
4:00 272.6 F 29.5  
  
Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 226.0 F 0.0  
10:00 230.4 F 0.0  
15:00 227.7 F 0.0  
20:00 225.1 F 0.0  
25:00 225.2 F 0.0  
30:00 224.6 F 0.0  
  
Drying Complete

FILLING: 2:30  
HEATING: 18:26  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

05/31/2023 03:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_