
Midmark M11 - v1.0.5
Total Cycles: 253

Name: RITTER2
Sterilizer ID: V2313228
07/27/2021 01:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.0 F 0.0
2:00 101.7 F 0.7
4:00 110.6 F 2.4
6:00 144.1 F 5.0
8:00 165.4 F 7.9
10:00 186.5 F 11.0
12:00 200.5 F 5.4
14:00 212.8 F 7.0
16:00 224.1 F 11.0
18:00 228.7 F 7.5
20:00 239.1 F 12.3
22:00 239.0 F 10.6
24:00 250.0 F 16.1
26:00 260.1 F 21.9
28:00 269.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.6 F 30.3
2:00 272.4 F 29.8
3:00 271.6 F 29.3
4:00 272.2 F 30.1

Min 271.5 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 209.0 F 0.0
10:00 214.7 F 0.0
15:00 208.8 F 0.0
20:00 222.3 F 0.0
25:00 218.3 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 29:07
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:07:47
CYCLE COMPLETE

07/28/2021 06:14 AM
ACCEPTED: X
REJECTED:
Operator: U01