
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000403
14:27 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:33
STERILIZING 3:35
136.0 C 220kPa 3:35
136.2 C 222kPa 4:05
136.0 C 218kPa 4:35
136.4 C 225kPa 5:05
136.2 C 220kPa 5:35
136.4 C 223kPa 6:05
136.4 C 225kPa 6:35
136.3 C 224kPa 7:05
136.3 C 224kPa 7:05
VENTING 7:05
AIR DRYING 7:55
CYCLE COMPLETE 8:00

Digital Signature #
3605975A8F8DD13C
