
Midmark M11 - v1.0.0
Total Cycles: 1198

Name: RITTER 1
Sterilizer ID: V2178198
09/23/2021 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.2 F 0.0
2:00 143.5 F 0.0
4:00 165.2 F 0.0
6:00 196.2 F 0.1
8:00 212.2 F 1.8
10:00 224.0 F 5.2
12:00 234.0 F 9.1
14:00 242.6 F 12.2
16:00 241.3 F 11.3
18:00 238.2 F 9.5
20:00 238.1 F 9.0
22:00 250.9 F 15.5
24:00 263.0 F 22.4
26:00 272.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.7 F 28.8
4:00 271.9 F 28.1

Min 271.5 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 233.6 F 0.0
10:00 202.9 F 0.0
15:00 201.7 F 0.0
20:00 202.7 F 0.0
25:00 203.2 F 0.0
30:00 210.7 F 0.0

Drying Complete

FILLING: 1:06
HEATING: 26:10
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

09/23/2021 09:47 AM
ACCEPTED: X
REJECTED:
Operator: U02