
Midmark M11 - v1.0.5
Total Cycles: 75

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.7 F 0.0
2:00 121.4 F 1.0
4:00 130.2 F 3.3
6:00 154.9 F 6.4
8:00 176.9 F 10.1
10:00 197.4 F 7.3
12:00 212.1 F 6.3
14:00 223.7 F 10.6
16:00 228.4 F 8.1
18:00 239.4 F 12.4
20:00 239.8 F 11.0
22:00 251.5 F 16.9
24:00 262.0 F 23.0
26:00 271.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 272.7 F 30.0
2:00 271.9 F 29.3
3:00 272.6 F 30.1
4:00 272.4 F 29.7

Min 271.6 F 29.2
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 201.4 F 0.0
10:00 212.2 F 0.0
15:00 214.7 F 0.0
20:00 214.1 F 0.0
25:00 204.7 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 26:36
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:55
CYCLE COMPLETE

06/08/2021 09:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____