
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005936
11:25 10/07/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:28
STERILIZING 3:20
136.0 C 220kPa 3:20
136.4 C 225kPa 3:50
136.4 C 225kPa 4:20
136.1 C 222kPa 4:50
136.3 C 224kPa 5:20
136.4 C 225kPa 5:50
136.4 C 224kPa 6:20
136.1 C 218kPa 6:50
136.1 C 218kPa 6:50
VENTING 6:50
AIR DRYING 7:39
CYCLE COMPLETE 7:44

Digital Signature #
8D5F2C2875F6A843
