
Midmark M11 - v1.0.5
Total Cycles: 2641

Name: RITTER2
Sterilizer ID: V2313228
07/13/2023 09:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.3 F 0.0
2:00 177.3 F 2.0
4:00 190.2 F 6.5
6:00 212.0 F 12.1
8:00 224.1 F 7.2
10:00 237.1 F 12.9
12:00 238.6 F 10.9
14:00 237.7 F 9.8
16:00 251.7 F 16.7
18:00 263.7 F 23.8
20:00 271.6 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.6 F 29.6
2:00 271.5 F 29.1
3:00 272.6 F 29.7
4:00 271.7 F 29.0

Min 271.5 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.5 F 0.0
5:00 231.8 F 0.0
10:00 242.5 F 0.0
15:00 231.1 F 0.0
20:00 226.7 F 0.0
25:00 221.7 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

07/13/2023 10:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____