

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2896

Name: RITTER2  
Sterilizer ID: V2313228  
09/18/2023 09:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 161.7 F 0.0  
2:00 162.4 F 1.6  
4:00 170.1 F 4.8  
6:00 191.6 F 9.0  
8:00 209.3 F 12.1  
10:00 224.3 F 8.5  
12:00 235.7 F 13.4  
14:00 237.1 F 10.5  
16:00 235.1 F 8.4  
18:00 248.8 F 14.9  
20:00 259.6 F 21.1  
22:00 269.8 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.4  
1:00 272.6 F 29.5  
2:00 272.7 F 29.6  
3:00 272.6 F 29.7  
4:00 272.6 F 29.9

Min 271.2 F 29.3  
Max 273.0 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.6 F 0.0  
5:00 244.6 F 0.0  
10:00 190.9 F 0.0  
15:00 187.6 F 0.0  
20:00 185.3 F 0.0  
25:00 183.2 F 0.0  
30:00 182.8 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 23:32  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:02:33  
CYCLE COMPLETE

09/18/2023 10:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_