

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005774  
9:02 19/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:18  
STERILIZING 4:22  
136.4 C 222kPa 4:22  
136.4 C 221kPa 4:52  
136.4 C 223kPa 5:22  
136.2 C 223kPa 5:52  
136.1 C 220kPa 6:22  
136.4 C 223kPa 6:52  
136.3 C 224kPa 7:22  
136.3 C 224kPa 7:52  
136.3 C 224kPa 7:52  
VENTING 7:52  
AIR DRYING 8:44  
CYCLE COMPLETE 19:43

-----  
Digital Signature #  
74DD0CE1CBE8B4B3  
-----