
Midmark M11 - v1.0.5
Total Cycles: 1066

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.8 F 0.0
2:00 171.1 F 3.4
4:00 189.0 F 8.2
6:00 209.7 F 10.4
8:00 230.5 F 9.0
10:00 236.0 F 10.2
12:00 242.8 F 12.6
14:00 242.2 F 11.7
16:00 255.7 F 18.7
18:00 267.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.6 F 29.9
2:00 271.7 F 29.2
3:00 272.7 F 29.5
4:00 272.6 F 29.9

Min 270.9 F 28.6
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 148.9 F 0.0
10:00 160.5 F 0.0
15:00 158.8 F 0.0
20:00 151.7 F 0.0
25:00 149.4 F 0.0
30:00 148.3 F 0.0

Drying Complete

FILLING: 1:50
HEATING: 19:16
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:20
CYCLE COMPLETE

05/03/2022 03:41 PM
ACCEPTED: X
REJECTED:
Operator: U02