

-----  
Midmark M11 - v1.0.5  
Total Cycles: 576

Name: RITTER2  
Sterilizer ID: V2313228  
11/24/2021 04:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.9 F 0.0  
2:00 193.6 F 3.5  
4:00 198.8 F 8.5  
6:00 214.6 F 13.3  
8:00 222.6 F 8.2  
10:00 236.8 F 13.0  
12:00 237.6 F 10.8  
14:00 237.2 F 9.3  
16:00 249.6 F 15.7  
18:00 260.6 F 21.8  
20:00 271.9 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.3 F 29.3  
2:00 271.4 F 29.3  
3:00 272.1 F 29.3  
4:00 271.1 F 29.1

Min 271.0 F 28.8  
Max 272.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.1 F 0.0  
5:00 210.0 F 0.0  
10:00 209.0 F 0.0  
15:00 203.6 F 0.0  
20:00 196.5 F 0.0  
25:00 191.8 F 0.0  
30:00 198.3 F 0.0  
Drying Complete

FILLING: 2:33  
HEATING: 20:31  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

11/24/2021 05:06 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_