
Midmark M11 - v1.0.0
Total Cycles: 3971

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2023 09:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 171.9 F 2.2
4:00 176.8 F 5.7
6:00 199.3 F 10.6
8:00 217.1 F 4.3
10:00 228.4 F 8.8
12:00 239.8 F 12.0
14:00 240.5 F 10.9
16:00 237.4 F 9.0
18:00 251.2 F 15.5
20:00 262.8 F 22.2
22:00 272.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.8 F 28.6
2:00 271.4 F 27.9
3:00 272.7 F 28.7
4:00 272.0 F 28.2

Min 271.4 F 27.9
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.2 F 0.0
5:00 241.8 F 0.0
10:00 217.7 F 0.0
15:00 224.8 F 0.0
20:00 228.9 F 0.0
25:00 230.7 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:16
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:01:12
CYCLE COMPLETE

11/02/2023 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____