

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3230

Name: RITTER2  
Sterilizer ID: V2313228  
12/18/2023 04:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 137.5 F 0.0  
2:00 137.2 F 1.8  
4:00 164.8 F 5.6  
6:00 192.1 F 10.5  
8:00 215.3 F 4.6  
10:00 230.9 F 10.0  
12:00 231.3 F 8.1  
14:00 242.7 F 12.7  
16:00 241.0 F 11.5  
18:00 254.7 F 18.5  
20:00 266.8 F 25.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 272.5 F 29.5  
2:00 272.4 F 29.9  
3:00 271.9 F 29.0  
4:00 272.9 F 29.8

Min 271.6 F 28.9  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
5:00 199.8 F 0.0  
10:00 201.9 F 0.0  
15:00 203.0 F 0.0  
20:00 202.1 F 0.0  
25:00 205.2 F 0.0  
30:00 211.1 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

12/18/2023 05:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_