

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001619  
12:23 11/25/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:07  
PRESSURIZING 6:53  
STERILIZING 8:36  
136.0 C 220kPa 8:36  
136.4 C 223kPa 9:06  
136.2 C 221kPa 9:36  
136.0 C 221kPa 10:06  
136.4 C 221kPa 10:36  
136.3 C 222kPa 11:06  
136.3 C 223kPa 11:36  
136.1 C 221kPa 12:06  
VENTING 12:06  
AIR DRYING 13:13  
CYCLE COMPLETE 23:46  
-----

Digital Signature #  
1AA8AB667D19D1B9  
-----