
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004198
8:11 16/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:00
STERILIZING 4:02
136.5 C 222kPa 4:02
136.3 C 220kPa 4:32
136.4 C 223kPa 5:02
136.3 C 221kPa 5:32
136.4 C 220kPa 6:02
136.4 C 222kPa 6:32
136.4 C 222kPa 7:02
136.3 C 223kPa 7:32
136.3 C 223kPa 7:32
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 14:39

Digital Signature #
9F636C328911EBAE
