
Midmark M11 - v1.0.5
Total Cycles: 120

Name: RITTER2
Sterilizer ID: V2313228
06/21/2021 07:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.2 F 0.0
2:00 73.6 F 0.1
4:00 79.9 F 0.8
6:00 102.0 F 2.8
8:00 135.2 F 5.4
10:00 164.1 F 8.3
12:00 179.7 F 12.0
14:00 200.7 F 3.7
16:00 213.4 F 7.0
18:00 224.3 F 10.7
20:00 229.8 F 8.2
22:00 237.8 F 11.6
24:00 236.5 F 9.7
26:00 248.8 F 15.6
28:00 259.7 F 21.5
30:00 269.3 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.5 F 29.9
2:00 271.6 F 29.7
3:00 272.2 F 29.8
4:00 271.7 F 29.8

Min 271.4 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 205.9 F 0.0
10:00 196.1 F 0.0
15:00 191.3 F 0.0
20:00 198.6 F 0.0
25:00 198.1 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

06/21/2021 08:59 AM
ACCEPTED: X
REJECTED:
Operator: U01