
Midmark M11 - v1.0.5
Total Cycles: 1945

Name: RITTER2
Sterilizer ID: V2313228
01/18/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.4 F 0.0
2:00 157.1 F 3.1
4:00 173.9 F 6.9
6:00 197.0 F 12.0
8:00 216.1 F 6.5
10:00 230.2 F 11.8
12:00 233.2 F 8.8
14:00 239.3 F 10.5
16:00 245.6 F 13.7
18:00 257.9 F 20.3
20:00 268.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 272.5 F 30.1
2:00 272.0 F 29.4
3:00 271.1 F 29.3
4:00 272.3 F 30.1

Min 271.0 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 225.3 F 0.0
10:00 235.1 F 0.0
15:00 231.7 F 0.0
20:00 227.3 F 0.0
25:00 229.6 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:10
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

01/18/2023 01:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____