

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4763

Name: RITTER 1  
Sterilizer ID: V2178198  
07/24/2024 01:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 86.7 F 0.0  
2:00 88.4 F 0.2  
4:00 103.1 F 1.3  
6:00 135.9 F 3.5  
8:00 165.0 F 6.2  
10:00 196.7 F 9.6  
12:00 207.6 F 13.5  
14:00 213.5 F 5.3  
16:00 223.9 F 9.1  
18:00 234.2 F 13.3  
20:00 233.6 F 9.5  
22:00 239.0 F 10.0  
24:00 243.9 F 12.6  
26:00 255.3 F 18.3  
28:00 265.4 F 24.4  
30:00 270.9 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.8 F 29.6  
2:00 271.7 F 29.0  
3:00 272.7 F 29.7  
4:00 271.9 F 29.0

Min 271.6 F 28.9  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.8 F 0.0  
5:00 206.2 F 0.0  
10:00 202.4 F 0.0  
15:00 200.8 F 0.0  
20:00 203.8 F 0.0  
25:00 211.2 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:47  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:09:23  
CYCLE COMPLETE

07/24/2024 02:41 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_