
Midmark M11 - v1.0.5
Total Cycles: 2172

Name: RITTER2
Sterilizer ID: V2313228
03/24/2023 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.1 F 0.0
2:00 83.0 F 0.0
4:00 121.2 F 0.0
6:00 115.6 F 0.1
8:00 147.8 F 2.6
10:00 176.4 F 5.7
12:00 200.7 F 9.2
14:00 211.2 F 8.9
16:00 221.9 F 6.9
18:00 233.0 F 11.2
20:00 232.0 F 7.9
22:00 244.4 F 13.0
24:00 243.5 F 12.4
26:00 255.6 F 18.7
28:00 266.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.1 F 29.0
2:00 272.6 F 29.4
3:00 272.0 F 29.4
4:00 271.8 F 28.8

Min 271.6 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.5 F 0.0
5:00 216.2 F 0.0
10:00 204.1 F 0.0
15:00 199.4 F 0.0
20:00 196.7 F 0.0
25:00 193.8 F 0.0
30:00 199.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 29:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

03/24/2023 09:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____