

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000415  
14:56 03/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:41  
STERILIZING 3:45  
136.0 C 221kPa 3:45  
136.2 C 223kPa 4:15  
136.4 C 224kPa 4:45  
136.1 C 222kPa 5:15  
136.2 C 223kPa 5:45  
136.2 C 223kPa 6:15  
136.4 C 225kPa 6:45  
136.2 C 219kPa 7:15  
136.2 C 219kPa 7:15  
VENTING 7:15  
AIR DRYING 8:09  
CYCLE COMPLETE 8:28

-----  
Digital Signature #  
C977CA03C90A9F6D  
-----