

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5216

Name: RITTER 1  
Sterilizer ID: V2178198  
01/02/2025 11:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.8 F 0.0  
2:00 133.8 F 0.0  
4:00 146.2 F 0.2  
6:00 167.5 F 0.9  
8:00 192.1 F 4.6  
10:00 208.1 F 8.5  
12:00 227.2 F 12.9  
14:00 229.5 F 8.1  
16:00 240.4 F 13.1  
18:00 239.1 F 10.5  
20:00 235.8 F 8.2  
22:00 249.4 F 14.9  
24:00 260.9 F 21.3  
26:00 271.2 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.4 F 28.8  
2:00 271.5 F 28.3  
3:00 272.4 F 28.8  
4:00 271.4 F 28.2

Min 271.3 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 207.4 F 0.0  
10:00 217.4 F 0.0  
15:00 229.6 F 0.0  
20:00 234.8 F 0.0  
25:00 229.9 F 0.0  
30:00 226.5 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 26:41  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:05:11  
CYCLE COMPLETE

01/02/2025 12:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_