
Midmark M11 - v1.0.0
Total Cycles: 1398

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2021 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.4 F 0.0
2:00 157.0 F 1.4
4:00 172.6 F 4.5
6:00 196.1 F 8.6
8:00 213.3 F 9.9
10:00 228.0 F 7.9
12:00 239.0 F 13.1
14:00 238.4 F 10.2
16:00 235.9 F 8.1
18:00 249.7 F 15.1
20:00 261.3 F 21.7
22:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.2 F 28.4
2:00 272.4 F 28.9
3:00 271.3 F 28.3
4:00 272.2 F 28.6

Min 271.3 F 28.2
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.2 F 0.0
5:00 208.4 F 0.0
10:00 209.0 F 0.0
15:00 230.2 F 0.0
20:00 221.5 F 0.0
25:00 220.2 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:27
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:00
CYCLE COMPLETE

11/29/2021 10:47 AM
ACCEPTED: X
REJECTED:
Operator: U02