
Midmark M11 - v1.0.0
Total Cycles: 1249

Name: RITTER 1
Sterilizer ID: V2178198
10/08/2021 11:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.1 F 0.0
2:00 122.6 F 0.0
4:00 155.4 F 0.1
6:00 182.8 F 1.4
8:00 201.8 F 5.2
10:00 215.4 F 9.2
12:00 227.7 F 10.6
14:00 233.9 F 9.0
16:00 239.9 F 10.8
18:00 243.7 F 12.0
20:00 241.1 F 10.6
22:00 255.0 F 17.5
24:00 266.6 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.6
2:00 271.4 F 27.8
3:00 272.7 F 28.7
4:00 271.6 F 27.9

Min 271.3 F 27.7
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 167.9 F 0.0
5:00 225.6 F 0.0
10:00 189.3 F 0.0
15:00 183.2 F 0.0
20:00 180.7 F 0.0
25:00 179.4 F 0.0
30:00 178.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:32
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:03:39
CYCLE COMPLETE

10/08/2021 01:00 PM
ACCEPTED: X
REJECTED:
Operator: U02