
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001208
14:17 18/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:26
PRESSURIZING 3:27
STERILIZING 4:17
136.0 C 220kPa 4:17
136.4 C 225kPa 4:47
136.0 C 221kPa 5:17
136.1 C 220kPa 5:47
136.4 C 224kPa 6:17
136.4 C 224kPa 6:47
136.4 C 223kPa 7:17
136.4 C 224kPa 7:47
136.4 C 224kPa 7:47
VENTING 7:47
AIR DRYING 8:44
CYCLE COMPLETE 8:57

Digital Signature #
04CB4E004B282A89
