
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001387
13:18 16/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:43
PRESSURIZING 2:46
STERILIZING 4:06
136.3 C 222kPa 4:06
136.4 C 225kPa 4:37
136.4 C 223kPa 5:07
136.4 C 223kPa 5:37
136.2 C 223kPa 6:07
136.3 C 223kPa 6:37
136.2 C 224kPa 7:07
136.4 C 225kPa 7:37
136.4 C 225kPa 7:37
VENTING 7:37
AIR DRYING 8:32
CYCLE COMPLETE 23:23

Digital Signature #
1BB926934BE1D577
