
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001674
13:51 04/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:53
STERILIZING 3:57
136.4 C 224kPa 3:57
136.2 C 223kPa 4:27
136.1 C 223kPa 4:57
136.4 C 225kPa 5:27
136.4 C 224kPa 5:57
136.3 C 224kPa 6:27
136.4 C 225kPa 6:57
136.4 C 224kPa 7:27
136.4 C 224kPa 7:27
VENTING 7:27
AIR DRYING 8:18
CYCLE COMPLETE 38:18

Digital Signature #
CB4365CC23D0CD15
