
Midmark M11 - v1.0.5
Total Cycles: 1628

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 08:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 158.4 F 2.3
4:00 175.4 F 5.7
6:00 197.8 F 10.4
8:00 220.4 F 5.2
10:00 232.7 F 10.2
12:00 231.4 F 7.6
14:00 243.6 F 12.9
16:00 241.8 F 11.6
18:00 254.6 F 18.2
20:00 266.1 F 25.2
22:00 272.7 F 30.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.7
1:00 271.5 F 29.7
2:00 271.2 F 29.6
3:00 271.7 F 29.6
4:00 272.2 F 29.7

Min 271.2 F 29.5
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 231.6 F 0.0
10:00 227.6 F 0.0
15:00 228.1 F 0.0
20:00 229.8 F 0.0
25:00 226.9 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:55
CYCLE COMPLETE

10/20/2022 09:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____