
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002625
14:51 27/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:57
STERILIZING 3:03
136.5 C 226kPa 3:03
136.2 C 224kPa 3:33
136.4 C 224kPa 4:03
136.1 C 223kPa 4:33
136.4 C 226kPa 5:03
136.4 C 226kPa 5:33
136.2 C 224kPa 6:03
136.2 C 224kPa 6:33
136.2 C 224kPa 6:33
VENTING 6:33
AIR DRYING 7:23
CYCLE COMPLETE 31:54

Digital Signature #
F9477BCA115E7438
