
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.4uS / 0.8ppm
CYCLE NUMBER 000622
13:24 01/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:22
136.0 C 228kPa 4:22
136.4 C 233kPa 4:52
136.2 C 231kPa 5:22
136.4 C 233kPa 5:52
136.2 C 231kPa 6:22
136.4 C 233kPa 6:52
136.2 C 232kPa 7:22
136.2 C 232kPa 7:52
136.2 C 232kPa 7:52
VENTING 7:52
AIR DRYING 8:47
CYCLE COMPLETE 23:47

Digital Signature #
57CA509A472B4124
