
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000798
16:07 16/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 4:46
STERILIZING 6:05
136.0 C 220kPa 6:05
135.9 C 219kPa 6:35
136.3 C 222kPa 7:05
136.4 C 218kPa 7:35
136.1 C 220kPa 8:05
136.4 C 221kPa 8:35
135.9 C 219kPa 9:05
136.4 C 223kPa 9:35
136.4 C 223kPa 9:35
VENTING 9:35
AIR DRYING 10:35
CYCLE COMPLETE 23:30

Digital Signature #
8775077C4F5A7E05
