
Midmark M11 - v1.0.0
Total Cycles: 1606

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 10:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 125.8 F 0.0
2:00 137.8 F 1.6
4:00 152.3 F 4.8
6:00 175.3 F 8.7
8:00 197.5 F 13.4
10:00 215.7 F 6.1
12:00 229.3 F 10.9
14:00 230.3 F 6.6
16:00 242.4 F 12.6
18:00 240.7 F 10.8
20:00 253.7 F 17.3
22:00 265.0 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.3 F 29.2
4:00 272.4 F 28.9

Min 271.5 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 221.8 F 0.0
10:00 227.4 F 0.0
15:00 229.8 F 0.0
20:00 230.1 F 0.0
25:00 224.7 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

02/08/2022 11:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____