
Midmark M11 - v1.0.5
Total Cycles: 814

Name: RITTER2
Sterilizer ID: V2313228
02/17/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.3 F 0.0
2:00 193.4 F 3.4
4:00 200.9 F 9.0
6:00 221.5 F 8.1
8:00 231.1 F 11.3
10:00 235.9 F 9.6
12:00 236.6 F 9.1
14:00 249.8 F 16.0
16:00 262.4 F 23.1
18:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.4 F 30.0
2:00 271.4 F 29.2
3:00 271.9 F 30.1
4:00 272.0 F 29.8

Min 271.2 F 29.2
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 232.0 F 0.0
10:00 229.4 F 0.0
15:00 227.9 F 0.0
20:00 223.3 F 0.0
25:00 217.6 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:03
CYCLE COMPLETE

02/17/2022 11:39 AM
ACCEPTED: X
REJECTED:
Operator: U02