
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076

Unit # : 000

WATER QUALITY

2.5uS / 1.6ppm

CYCLE NUMBER 000046

6:04 25/08/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 4:24

PRESSURIZING 10:13

STERILIZING 11:08

123.0 C 117kPa 11:08

123.4 C 120kPa 11:38

123.3 C 120kPa 12:08

123.4 C 120kPa 12:38

123.4 C 120kPa 13:08

123.4 C 120kPa 13:38

123.4 C 120kPa 14:08

123.4 C 120kPa 14:38

123.1 C 119kPa 15:09

123.3 C 120kPa 15:39

123.4 C 118kPa 16:09

123.4 C 120kPa 16:39

123.3 C 120kPa 17:09

123.5 C 120kPa 17:39

123.4 C 120kPa 18:09

123.5 C 120kPa 18:39

123.5 C 120kPa 19:09

123.4 C 120kPa 19:39

123.5 C 120kPa 20:09

123.4 C 120kPa 20:39

123.2 C 120kPa 21:09

123.2 C 120kPa 21:39

123.4 C 120kPa 22:09

123.4 C 120kPa 22:39

123.4 C 120kPa 23:09

123.5 C 120kPa 23:39

123.5 C 120kPa 24:09

123.4 C 120kPa 24:40

123.2 C 119kPa 25:10

123.5 C 120kPa 25:40

123.5 C 120kPa 26:08

VENTING 26:08

AIR DRYING 27:08
CYCLE COMPLETE 57:08

Digital Signature #
AD9570CAAA3A71C0
