

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2691

Name: RITTER2  
Sterilizer ID: V2313228  
07/24/2023 05:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.9 F 0.0  
2:00 160.7 F 0.4  
4:00 175.7 F 4.0  
6:00 210.5 F 8.0  
8:00 225.0 F 12.0  
10:00 224.8 F 6.3  
12:00 233.5 F 10.6  
14:00 235.3 F 7.8  
16:00 242.3 F 12.9  
18:00 242.1 F 11.8  
20:00 253.8 F 17.8  
22:00 264.6 F 24.3  
24:00 271.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.9  
1:00 272.5 F 29.7  
2:00 271.8 F 29.1  
3:00 271.2 F 29.1  
4:00 272.4 F 29.7

Min 271.1 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 216.6 F 0.0  
10:00 223.9 F 0.0  
15:00 228.3 F 0.0  
20:00 222.2 F 0.0  
25:00 217.5 F 0.0  
30:00 216.7 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 24:06  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:02:45  
CYCLE COMPLETE

07/24/2023 06:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_