
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002479
12:12 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:53
STERILIZING 3:58
136.0 C 220kPa 3:58
136.3 C 225kPa 4:28
136.2 C 223kPa 4:58
136.1 C 222kPa 5:28
136.3 C 222kPa 5:58
136.2 C 222kPa 6:28
136.1 C 223kPa 6:58
136.2 C 224kPa 7:28
136.2 C 224kPa 7:28
VENTING 7:28
AIR DRYING 8:18
CYCLE COMPLETE 15:38

Digital Signature #
95BC09BC35F14E5C
