
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000057
8:35 26/06/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:52
STERILIZING 4:05
136.0 C 220kPa 4:05
136.4 C 223kPa 4:35
136.3 C 223kPa 5:05
136.2 C 222kPa 5:35
136.1 C 222kPa 6:05
136.4 C 224kPa 6:35
136.3 C 221kPa 7:05
136.5 C 225kPa 7:35
136.5 C 225kPa 7:35
VENTING 7:35
AIR DRYING 8:26
CYCLE COMPLETE 23:26

Digital Signature #
EBCBCD086B1A80AC
